

Model 15i HCl Analyzer

Gas filter correlation gas analyzer with enhanced communication capabilities for source emissions monitoring



Key Features

- ◆ Ethernet connectivity for efficient remote access
- ◆ Enhanced user interface with one button programming and large display screen
- ◆ Flash memory for increased data storage and user downloadable software
- ◆ Enhanced electronics design optimizes product commonality
- ◆ Improved layout for easier accessibility to components

A core product in the Thermo Scientific *iSeries* product line is the Model 15*i* analyzer. Using gas filter correlation technology, the Model 15*i* measures the amount of Hydrogen Chloride in the sample. The Model 15*i* is based on the principle that Hydrogen Chloride (HCl) absorbs infrared radiation at a wavelength of 3.4 microns. Because infrared absorption is a nonlinear measurement technique, it is necessary for the instrument electronics to transform the basic analyzer signal into a linear output.

The Model 15*i* uses an exact calibration curve to accurately linearize the instrument output over any range up to a concentration of 5,000ppm.

This state-of-the-art gas analyzer offers features such as an ethernet port, flash memory for increased data storage as well as being field upgradable.

Ethernet connectivity provides efficient remote access, allowing the user to download measurement information directly from the instrument without having to be on-site.

Easily programmable shortcut-keys allow you to jump directly to frequently accessed functions, menus or screens. The larger interface screen can display measurement information and status while viewing menu and operational screens.

Comprehensive Service Solutions

To maintain optimal product performance, you need immediate access to experts worldwide, as well as priority status when your air quality equipment needs repair or replacement. We offer comprehensive, flexible support solutions for all phases of the product lifecycle. Through predictable, fixed-cost pricing, our services help protect the return on investment and total cost of ownership of your Thermo Scientific air quality products.

Product Specifications

Preset Ranges	0-5, 10, 20, 50, 100, 200, 500, 1000, 2000, and 5000 ppm 0-10, 20, 50, 100, 200, 500, 1000, 2000, 5000, and 7500 mg/m ³
Custom Ranges	0-5 to 5000 ppm 0-10 to 7500 mg/m ³
Zero Noise	0.1 ppm RMS (5 minute averaging time)
Lower Detectable Limit	0.20 ppm
Zero Drift (24 hour)	+/- 0.2 ppm
Span Drift (24 hour)	+/-2% full scale
Response Time	120 seconds (30 second average time)
Precision	+/-0.1 ppm
Linearity	+/-2% full scale < 1000 ppm, +/- 5% full scale >1000 ppm
Sample Flow Rate	1.0 liters/min. nominal
Operating Temperature	5°C - 45°C
Power Requirements	100 vac, 115 vac, 220-240 vac +/-10% @ 275W
Size and Weight	16.75"(W) x 8.62"(H) x 23"(D), 49 lbs. 425 mm (W) x 219 mm (H) x 584 mm (D), 22.2 kg
Outputs	Selectable Voltage, RS232/RS485, TCP/IP, 10 Status Relays, and Power Fail Indication (standard). 0-20 or 4-20 mA Isolated Current Output (optional)
Inputs	16 Digital Inputs (standard), 8 0-10Vdc Analog Inputs (optional)

Ordering Information

Model 15i-HCl Analyzer

Choose from the following configurations/options to customize your own Model 15i

Voltage options:

A = 115 Vac 50/60 Hz
B = 220/240 Vac 50/60 Hz
J = 100 Vac 50/60 Hz

Internal zero /span:

N = No zero / span valve
Z = Internal zero / span valves

Filter Wheel Purge:

S = Standard plumbing (standard)
P = Filter wheel purge set-up

Optional I/O:

A = No optional I/O (standard)
C = 0-20, 4-20mA current output - 6 channels, 0-10v analog input - 8 channels

Mounting hardware:

A = Bench mounting (standard)
B = Ears & handles, EIA
C = Ears & handles, retrofit

Your Order Code: 15i - _ _ _ _ _



Other options:

- Teflon particulate filter
- Rack mounts
- Rear extender
- Terminal Block Kit & Cable 37 pin
- Terminal Block Kit & Cable 25 pin
- Cable, DB37M to open end, 6' LG.
- Cable, DB37F to open end, 6' LG.
- Cable, DB25M to open end, 6' LG.
- Cable, RS232 Null Modem

This specification sheet is for informational purposes only and is subject to change without notice. Thermo makes no warranties, expressed or implied, in this product summary.
© 2007 Thermo Fisher Scientific Inc. All rights reserved Thermo Fisher Scientific Inc.

Environmental
Instruments
Air Quality Instruments

27 Forge Parkway
Franklin, MA
02038 USA

+1 (866) 282-0430
+1 (508) 520-1460 fax

www.thermo.com/air

Takkebijsters 1
Breda, 4817 BL
The Netherlands

+33 (0) 76 5879888
+33 (0) 76 5816520 fax

Lit_15iAQI_3/08

Thermo
SCIENTIFIC